

Viper Test Strips Directions

Directions for Use:

Dip the strip in the test solution for one second. Gently shake off the excess fluid. At exactly 120 seconds, compare the strip color with the color block. The test strip should show a reading of the color that is circled, 1000 ppm. The strip color will continue to develop after the read time. Delayed reading of the strip color may invalidate the results.

Please let Cherie Hardie know if you require additional test strips - chardie@hpedsb.on.ca or 613-966-1170 ext. 2377

